



Quality Certificate

Product Name: **Easy Flip™ 15mL Centrifuge Tubes**

Catalog Number : CFT211150

Customer Focus:

At Jet, our customers' satisfaction is exactly what we are pursuing. We have good management and product development teams who have years of industry experience. Our cell and tissue culture professionals and our entrepreneurial spirit make us responsive to customer needs. We welcome special requests and work to provide the best and most personal service in the industry.

Satisfaction Guarantee:

Jet offers a total satisfaction guarantee so you can be confident in your purchasing decision. If for any reason you are not satisfied with the product performance or service provided by Jet we will either replace or issue a refund for the purchase price of your product.

JET products are manufactured in accordance with ISO13485:2003&ISO 9001:2008 quality management systems in a class 100,000 clean room.

Sterilization - Product labeled as sterile is irradiated and dosimetrically released upon ISO 11137 recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. JET BIOFIL products labeled sterile meet a minimum requirement of 10^{-6} SAL (Sterility Assurance Level).

Non-Pyrogenic - Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate) testing for medical devices and company guidelines. The acceptance level for product is less than 0.5 EU/mL.

DNase&RNase Free - This product has been tested and is free of any detectable RNase/DNase contamination.

Max. Relative Centrifuge force - Products labeled Max. Relative Centrifuge force has been validated on centrifuge force testing. The acceptance level for product with Conical bottom is $9,400\times g$.

Autoclave Test - This product has been tested at 121°C for 30min in autoclave and it is integrate, no distortion and no leak with cap after autoclave.

Low temperature test - This product has been tested at -80°C for 48hour in freezer and it is integrate, no distortion and no leak with cap after freezing.

Quality Control Testing - This product has been inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

YORK ZHANG
Quality Manager